

ANNEX E AIR QUALITY



CALIBRATION CERTIFICATES FOR AIR QUALITY

ALS Technichem (HK) Pty Ltd

ALS Laboratory Group





SUB-CONTRACTING REPORT

CONTACT

: MR MAGNUM FAN

WORK ORDER

HK2342916

CLIENT ADDRESS

PROJECT

: ENVIROTECH SERVICES CO.

: RM 712, 7/F, MY LOFT 9 HOI WING ROAD,

TUEN MUN, N.T. HK

SUB-BATCH

DATE RECEIVED : 26-OCT-2023

DATE OF ISSUE : 2-NOV-2023

NO. OF SAMPLES : 1

CLIENT ORDER

General Comments

Sample(s) was/ were submitted by client. Sample(s) arrived laboratory in ambient condition.

Sample information (Project name, Sample ID, Sampling date/time, etc.) is provided by client.

Result(s) of sample(s) is/are reported on as received basis, unless otherwise specified. The result(s) is/are related only to the item(s) tested.

Calibration was subcontracted to Envirotech Services Company.

Signatories

This document has been signed by those names that appear on this report and are the authorised signatories

Signatories

Position

Richard Fung

Managing Director

WORK ORDER

CLIENT

: HK2342916

SUB-BATCH

1 : ENVIROTECH SERVICES CO.

PROJECT



ALS Lab	Client's Sample ID	Sample Type	Sample Date	External Lab Report No.
HK2342916-001	Sibata LD-3B (436553)	Equipments	21-Oct-2023	S/N: 436553



Envirotech Services Co.

Rm. 712, 7/F My Loft, 9 Hoi Wing Road, Tuen Mun. H.K. Tel: 2560 8450 Fax: 2560 6553

E-mail: envirotech@netvigetor.com

Equipment Verification Report (TSP)

Easti	pment	Calib	rated.
CUUI	Dillell	Callin	iaccu.

Type:

Laser Dust Monitor

Manufacturer:

Sibata LD-3B

Serial No.:

436553

Equipment Ref.:

N/A

ALS Job Order:

HK2342309

Standard Equipment

Standard Equipment:

High Volume Sampler (TSP)

Location:

Envirotech Room (Calibration Room)

Equipment Ref.:

HVS 8162

Last Calibration Date:

21-Oct-2023

Equipment Verification Results:

Verification Date:

14-Oct-2023

		Mean	Mean	Concentration in µg/m³	Concentration in µg/m ³
Hour	Time	Temp °C	Pressure	(Standard Equipment)	(Calibrated Equipment)
			(hpa)	(Y-Axis)	(X-Axis)
1hr 00mins	0910-1010	21.8	1019	10	17
2hr 00mins	1015-1215	21.8	1019	30	38
3hr 00mins	. 1410-1710	23.5	1019	71	67

80

Linear Regression of Y or X


Slope (K-factor):

 $1.0622(\mu g/m^3)/CPM$

Correlation Coefficient (R):

Date of Issue:

0.9872 26-Oct-2023



Remarks:

- 1. Strong Correlation (>0.8)
- Factor 1.0622 (µg/m³)/CPM should be applied for TSP monitoring

Operator:

P.F.Yeung Signature

Da

Date: 26 October 2023

QC Reviewer:

K.F.Ho

Signature

Fail

Date: 26 October 2023

^{*}If R<0.5, repair or verification is required for the equipment

TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location: Rm. 712, My Loft, Tuen Mun Date of Calibration: 13-Oct-23 HVS ID: 8162 Next Calibration Date: 12-Dec-23 Name and Model: TISCH HVS Model TE-5170 Operator: P.F. Yeung CONDITIONS Sea Level Pressure (hpa) 1015 Corrected Pressure (mm Hg) 762.1 28.9 Temperature (°C) Temperature (K) 293 CALIBRATION ORIFICE Make: TISCH **Qstd Slope** 2.06918 Model: TE-5025A **Qstd** Intercept -0.04220Serial#: 2454 CALIBRATION Plate H2O(L) H20(R) H₂O IC I Qstd LINEAR No. (in) (in) (m3/min) (in) (chart) (corrected) REGRESSION 18 6.5 6.5 13.0 1.806 62 63.54 Slope= 32.843 13 4.7 4.7 9.4 1.539 56 57.39 Intercept= 5.518 10 3.4 3.4 6.8 1.312 49 50.22 Corr. Coeff.= 0.9939 7 2.3 2.2 4.5 1.071 40 40.99 5 1.6 1.5 3.1 0.892 33 33.82 IC Flow Rate Calulations: 70 Qstd = 1/m[Sqrt(H2O(Pa/Pstd)(Tstd/Ta))-b]65 IC = I[Sqrt(Pa/Pstd)(Tstd/Ta)]60 Qstd = standard flow rate 55 IC = corrected chart response 50 I = actual chart response m = calibrator Qstd slope 45 b = calibrator Qstd intercept 40 Ta = actual temperature during calibration (deg K)

35

30

25

20 15

10

0.7 0.8 0.9 1.0 1.1 1.2 1.3 1.4 1.5 1.6 1.7 1.8 1.9

Qstd(m3/min)

Pa = actual pressure during calibration (mm Hg)

For subsequent calculation of sampler flow:

1/m((I)[Sqrt(298/Tav)(Pav/760)]-b)

Tav = daily average temperature
Pav = daily average pressure

m = sampler slope
b = sampler intercept

I = chart response



RECALIBRATION DUE DATE:

December 15, 2023

Certificate of Calibration

Calibration Certification Information

Cal. Date: December 15, 2022

Rootsmeter S/N: 438320

Ta: 295

Pa: 748.0

°K

Operator: Jim Tisch

Calibrator S/N: 4064

mm Hg

Calibration Model #: TE-5025A

Run	Vol. Init (m3)	Vol. Final (m3)	ΔVol. (m3)	ΔTime (min)	ΔP (mm Hg)	ΔH (in H2O)
1	1	2	1	1.4430	3.2	2.00
2	3	4	1	1.0210	6.4	4.00
3	5	6	1	0.9170	7.9	5.00
4	7	8	1	0.8730	8.8	5.50
	0	10	1	0.7210	12.8	8.00

	Data Tabulation						
Vstd	Qstd	$\sqrt{\Delta H \left(\frac{Pa}{Pstd}\right) \left(\frac{Tstd}{Ta}\right)}$		Qa	√∆H(Ta/Pa)		
(m3)	(x-axis)	(y-axis)	Va	(x-axis)	(y-axis)		
0.9900	0.6861	1.4101	0.9957	0.6900	0.8881		
0.9858	0.9655	1.9943	0.9914	0.9711	1.2560		
0.9838	1.0728	2.2296	0.9894	1.0790	1.4042		
0.9826	1.1255	2.3385	0.9882	1.1320	1.4728		
0.9772	1.3554	2.8203	0.9829	1.3632	1.7762		
	m=	2.10977		m=	1.32110		
QSTD	b=	-0.03782	QA	b=	-0.02382		
2010	r=	0.99998		r=	0.99998		

	Calculation	ns	
Vstd=	ΔVol((Pa-ΔP)/Pstd)(Tstd/Ta)	Va=	ΔVol((Pa-ΔP)/Pa)
Qstd=	Vstd/ΔTime	Qa= Va/ΔTime	
	For subsequent flow rat	e calculatio	ns:
Qstd=	$1/m\left(\left(\sqrt{\Delta H\left(\frac{Pa}{Pstd}\right)\left(\frac{Tstd}{Ta}\right)}\right)-b\right)$	Qa=	$1/m\left(\left(\sqrt{\Delta H\left(Ta/Pa\right)}\right)-b\right)$

	Standard Conditions
Tstd:	298.15 °K
Pstd:	760 mm Hg
	Key
ΔH: calibrator	manometer reading (in H2O)
ΔP: rootsmete	er manometer reading (mm Hg)
Ta: actual abs	olute temperature (°K)
Pa: actual bar	ometric pressure (mm Hg)
b: intercept	
m: slope	

RECALIBRATION

US EPA recommends annual recalibration per 1998
40 Code of Federal Regulations Part 50 to 51,
Appendix B to Part 50, Reference Method for the
Determination of Suspended Particulate Matter in
the Atmosphere, 9.2.17, page 30

Tisch Environmental, Inc. 145 South Miami Avenue Village of Cleves, OH 45002 www.tisch-env.con

TOLL FREE: (877)263-7610 FAX: (513)467-900!



MONITORING SCHEDULE FOR AIR QUALITY

Tung Chung New Town Extension (East)
Air Quality Monitoring Schedule (May 2024)

Sunday	Monday	Tuesday	Wednesday	Thursday	Friday	Saturday
			1-May	2-May	3-May	4-May
				Air Quality Monitoring		
5-May	6-May	7-May	8-May	9-May	10-May	11-May
			Air Quality Monitoring			
12-May	13-May	14-May	15-May	16-May	17-May	18-May
1 - 1111	,	Air Quality Monitoring				
19-May	20-May	21-May	22-May	23-May	24-May	25-May
	Air Quality Monitoring					Air Quality Monitoring
26-May	27-May	28-May	29-May	30-May	31-May	
					Air Quality Monitoring	




MONITORING RESULTS FOR AIR QUALITY

Table E3 Data for 1-hr TSP Monitoring at Station DM-1

Date	Start Time	Finish Time	Weather	1-hour TSP (μg/m³)
5/2/2024	10:02	11:02	Cloudy	57
5/2/2024	9:02	10:02	Cloudy	54
5/2/2024	11:02	12:02	Cloudy	53
5/8/2024	9:00	10:00	Cloudy	60
5/8/2024	10:00	11:00	Cloudy	62
5/8/2024	11:00	12:00	Cloudy	61
5/14/2024	13:45	14:45	Sunny	37
5/14/2024	14:45	15:45	Sunny	41
5/14/2024	15:45	16:45	Sunny	49
5/20/2024	9:02	10:02	Cloudy	44
5/20/2024	10:02	11:02	Cloudy	46
5/20/2024	11:02	12:02	Cloudy	47
5/25/2024	13:08	14:08	Cloudy	46
5/25/2024	14:08	15:08	Cloudy	48
5/25/2024	15:08	16:08	Cloudy	51
5/31/2024	13:11	14:11	Cloudy	47
5/31/2024	14:11	15:11	Cloudy	49
5/31/2024	15:11	16:11	Cloudy	46

Figure E3 Graphical Presentation for 1-hr TSP Monitoring at Station DM-1





EVENT AND ACTION PLAN FOR AIR QUALITY

Annex E4 Event and Action Plan for Air Quality

Event		Action	1	
Event	ET	IEC	ER	Contractor
Action level exceedance for one sample	 Identify source, investigate the causes of exceedance and propose remedial measures; Inform IEC and ER; Repeat measurement to confirm finding; Increase monitoring frequency to daily. 	 Check monitoring data submitted by ET; Check Contractor's working method. 	1. Notify Contractor.	 Rectify any unacceptable practice; Amend working methods if appropriate.
Action level exceedance for two or more consecutive samples	 Identify source; Inform IEC and ER; Advise the ER on the effectiveness of the proposed remedial measures; Repeat measurements to confirm findings; Increase monitoring frequency to daily; Discuss with IEC and Contractor on remedial actions required; If exceedance continues, arrange meeting with IEC and ER; If exceedance stops, cease additional monitoring. 	 Check monitoring data submitted by ET; Check Contractor's working method; Discuss with ET and Contractor on possible remedial measures; Advise the ET on the effectiveness of the proposed remedial measures; Supervise Implementation of remedial measures. 	failure in writing;2. Notify Contractor;3. Ensure remedial measures properly implemented.	 Submit proposals for remedial to ER within 3 working days of notification; Implement the agreed proposals; Amend proposal if appropriate.

Frank	Action							
Event	ET	IEC	ER	Contractor				
Limit level exceedance for one sample	 Identify source, investigate the causes of exceedance and propose remedial measures; Inform ER, Contractor and EPD; Repeat measurement to confirm finding; Increase monitoring frequency to daily; Assess effectiveness of Contractor's remedial actions and keep IEC, EPD and ER informed of the results. 	 Check monitoring data submitted by ET; Check Contractor's working method; Discuss with ET and Contractor on possible remedial measures; Advise the ER on the effectiveness of the proposed remedial measures; Supervise implementation of remedial measures. 	failure in writing;Notify Contractor;Ensure remedial measures properly implemented.	 Take immediate action to avoid further exceedance; Submit proposals for remedial actions to IEC within 3 working days of notification; Implement the agreed proposals; Amend proposal if appropriate. 				
Limit level exceedance for two or more consecutive samples	 Notify IEC, ER, Contractor and EPD; Identify source; Repeat measurement to confirm findings; Increase monitoring frequency to daily; Carry out analysis of Contractor's working procedures to determine possible mitigation to be implemented; Arrange meeting with IEC and ER to discuss the remedial actions to be taken; Assess effectiveness of Contractor's remedial actions and keep IEC, EPD and ER informed of the results; If exceedance stops, cease additional monitoring. 	 Discuss amongst ER, ET, and Contractor on the potential remedial actions; Review Contractor's remedial actions whenever necessary to assure their effectiveness and advise the ER accordingly; Supervise the implementation of remedial measures. 	 Confirm receipt of notification of failure in writing; Notify Contractor; In consultation with the IEC, agree with the Contractor on the remedial measures to be implemented; Ensure remedial measures properly implemented; If exceedance continues, consider what portion of the work is responsible and instruct the Contractor to stop that portion of work until the exceedance is abated. 	 Take immediate action to avoid further exceedance; Submit proposals for remedial actions to IEC within 3 working days of notification; Implement the agreed proposals; Resubmit proposals if problem still not under control; Stop the relevant portion of works as determined by the ER until the exceedance is abated. 				